

EC DECLARATION OF CONFORMITY

We,

DIAsource ImmunoAssays S.A.

Rue du Bosquet, 2, B-1348 Louvain-la-Neuve, Belgium,

declare under our sole responsibility that the product

ELISA PLATE SHAKER (Cat. Nr. DIA4000)

Classification: Laboratory Equipment

The product described above is in conformity with the latest version of the following standards:

EN 61010-1 Safety requirements for electrical equipment
for measurement, control and laboratory use.

EN 61326-1 Part 1: General requirements. Electrical
equipment for measurement, control and
laboratory use – EMC requirements.

and the Directives and Annex:

2014/35/EU Low Voltage Directive (LVD)

2014/30/EU Electromagnetic Compatibility (EMC)

This product is manufactured in a QMS certified environment.

Technical documentation demonstrating compliance is kept by the manufacturer.

Louvain-la-Neuve, 29 September 2023

Béatrice de Borman
CEO & QRA Manager



DIASOURCE ImmunoAssays S.A.
Rue du Bosquet, 2
B-1348 Louvain-la-Neuve
RPM T.V.A. BE 0457 934 723
Tél.: +32 (0)10 84 99 22
Fax: +32 (0)10 84 99 90